

Proficiency Tests

DLA

food
cosmetics
consumer goods
www.dla-lvu.de

Evaluation Report
proficiency test

DLA 32/2017

Contaminated Food:
Heavy Metals (Pb, Cd, Hg, As)
in Root Vegetable Powder

Dienstleistung Lebensmittel Analytik GbR
Waldemar-Bonsels-Weg 170
22926 Ahrensburg, Germany

proficiency-testing@dl-lvu.de
www.dla-lvu.de

Coordinator: Dr. Gerhard Wichmann

Allgemeine Informationen zur Eignungsprüfung (EP)
General Information on the proficiency test (PT)

<i>EP-Anbieter</i> <i>PT-Provider</i>	DLA - Dienstleistung Lebensmittel Analytik GbR Gesellschafter: Dr. Gerhard Wichmann und Dr. Matthias Besler Waldemar-Bonsels-Weg 170, 22926 Ahrensburg, Germany Tel. ++49(0)171-1954375 Fax. ++49(0)4102-9944976 eMail. proficiency-testing@dla-lvu.de
<i>EP-Nummer</i> <i>PT-Number</i>	DLA 32/2017
<i>EP-Koordinator</i> <i>PT-Coordinator</i>	Dr. Gerhard Wichmann
<i>Status des EP-Bericht</i> <i>Status of PT-Report</i>	Abschlussbericht / Final report : 16 December 2017
<i>EP-Bericht Freigabe</i> <i>PT-Report Authorization</i>	Dr. Matthias Besler (Technischer Leiter / Technical Manager) - <i>gezeichnet / signed M. Besler</i> Dr. Gerhard Wichmann (QM-Beauftragter / Quality Manager) - <i>gezeichnet / signed G. Wichmann</i> Datum / Date: 16 December 2017
<i>Unteraufträge</i> <i>Subcontractors</i>	Die Prüfung der Gehalte, Homogenität und Stabilität von EP-Parametern wird von DLA im Unterauftrag vergeben. The analysis of the content, homogeneity and stability of PT-parameters are subcontracted by DLA.
<i>Vertraulichkeit</i> <i>Confidentiality</i>	Die Teilnehmerergebnisse sind im EP-Bericht in anonymisierter Form mit Auswertenummern benannt. Daten einzelner Teilnehmer werden ausschließlich nach vorheriger Zustimmung des Teilnehmers an Dritte weitergegeben. Participant result are named anonymously with evaluation numbers in the PT report. Data of individual participants will be passed on to third parties only with prior consent of the participant.

Inhalt / Content

1. Introduction.....	4
2. Realisation.....	4
2.1 Test material.....	4
2.1.1 Homogeneity.....	5
2.1.2 Stability.....	6
2.2 Sample shipment and information to the test.....	6
2.3 Results.....	6
3. Evaluation.....	7
3.1 Consensus values from participants (Assigned value).....	7
3.2 Robust standard deviation.....	7
3.3 Repeatability standard deviation.....	7
3.4 Reproducibility standard deviation.....	8
3.5 Exclusion of results and outliers.....	8
3.6 Target standard deviation (for proficiency assessment).....	9
3.6.1 General model (Horwitz).....	9
3.6.2 Precision experiment.....	10
3.6.3 Value by perception.....	11
3.7 z-Score.....	12
3.7.1 Warning and action signals.....	12
3.8 z'-Score.....	12
3.9 Reproducibility coefficient of variation (CV).....	13
3.10 Quotient S^*/opt	13
3.11 Standard uncertainty.....	14
4. Results.....	15
4.1 Lead (mg/kg).....	16
4.2 Cadmium (mg/kg).....	19
4.3 Arsenic (mg/kg).....	22
4.4 Mercury (mg/kg).....	25
5. Documentation.....	28
5.1 Details by participants.....	28
5.1.1 Primary data.....	28
5.1.2 Analytical methods.....	32
5.2 Homogeneity.....	36
5.2.1 Homogeneity testing before PT.....	36
5.2.2 Comparison of sample number/test results and trend line.....	37
5.3 Sample cover letter: Information on the Proficiency Test (PT)...	38
6. Index of participant laboratories.....	39
7. Index of literature.....	40

1. Introduction

The participation in proficiency testing schemes is an essential element of the quality-management-system of every laboratory testing food and feed, cosmetics and food contact materials. The implementation of proficiency tests enables the participating laboratories to prove their own analytical competence under realistic conditions. At the same time they receive valuable data regarding the verification and/or validation of the particular testing method [1, 5].

The purpose of DLA is to offer proficiency tests for selected parameters in concentrations with practical relevance.

Realisation and evaluation of the present proficiency test follows the technical requirements of DIN EN ISO/IEC 17043 (2010) and DIN ISO 13528:2009 / ISO 13528:2015 [2, 3].

2. Realisation

2.1 Test material

The test material was a plant-powder mixture (Carrot powder, winter cherry root powder, spirulina thallus powder) with a natural content of cadmium (Cd), lead (Pb), mercury (Hg) and arsenic (As) and a microtracer premix (wheat flour, microtracer iron particles (FSS red lake) for homogeneity verification.

Approximately 0,6 kg of the material were homogenized and sieved and then packaged in portions to approximately 5 g. The portions were numbered chronologically.

The detectability of the heavy metals ((Cd, Pb, Hg, As) was assured.

Note: The metrological traceability of temperature, mass and volume during production of the PT samples is ensured by DAkkS calibrated reference materials.

2.1.1 Homogeneity

The **mixture homogeneity before bottling** was examined 10-fold by **Tracer analysis**. It is a standardized method that is part of the international GMP certification system for feed [14].

Before mixing, fennel seeds are added to the sample and the number of particles is determined after homogenization in taken aliquots. The evaluation of the mixture homogeneity is based on the Poisson distribution using the chi-square test and based on the normal distribution using the HorRat value. For the evaluation according to Poisson: A probability of $\geq 5\%$ is equivalent to a good homogeneous mixture and of $\geq 25\%$ to an excellent mixture [14, 15]. For the evaluation according to the normal distribution: According to [16, 17], the HorRat values between 0,3 and 1,3 are to be accepted under repeatability conditions (measurements within the laboratory).

The tracer analysis of the present PT sample showed probability of 95%. Additionally particle number results were converted into concentrations, statistically evaluated according to normal distribution and compared to the standard deviation according to Horwitz. This gave a HorRat value of 0,6. The results of tracer analysis are given in the documentation.

The calculation of the variation coefficient (CV) of the repeatability standard deviation (CV_r) and of the reproducibility standard deviation (CV_R) was used as an indicator of homogeneity, see table 1. The coefficients (CV_r and CV_R) are comparable to the precision data of the official method, see 3.6.2/ table 2 [18-20, 22]. The repeatability standard deviations and the reproducibility standard deviations of the participants are given at the characteristics (4.1 to 4.4).

Tabelle 1: Compilation of the coefficients of variation CV_r and CV_R of the present PT.

	CV_r	CV_R
Lead	4,43%	9,16%
Cadmium	4,78%	16,0%
Arsenic	2,67%	6,53%
Mercury	11,3%	16,8%

Furthermore, the homogeneity was characterized by the **trend line function of participants' results for chronological bottled single samples**. The maximum deviations for cadmium from the mean value of the trend line was in the range of 35% of the target standard deviation σ_{pt} (s. 5.2 homogeneity) and is to be judged as acceptable.

If the criteria for sufficient homogeneity of the test material are not fulfilled on a particular parameter, the impact on the target standard deviation is checked and optionally the evaluation of the results of the participants will be done using the z'-score considering the standard uncertainty of the assigned value (see 3.8 and 3.11) [3].

2.1.2 Stability

The experience with various DLA reference materials showed good storage stability with respect to the durability of the sample (spoilage) and the content of lead, cadmium, arsenic and mercury for samples with a comparable water activity (a_w value <0.5) and matrix. The sample material is therefore stable against microbial spoilage at room temperature and dry light-protected storage.

2.2 Sample shipment and information to the test

Two portions of test material were sent to every participating laboratory in the 34th week of 2017. The testing method was optional. The tests should be finished at October 27th 2017 the latest.

With the cover letter along with the sample shipment the following information was given to participants:

In general we recommend to homogenize a representative sample amount before analysis according to good laboratory practice, especially in case of low sample weights.

Further information see 5.3.

2.3 Results

The participants submitted their results in standard forms, which have been handed out with the samples (by email).

The finally calculated concentrations as average of duplicate determinations of both numbered samples was used for the statistical evaluation. For the calculation of the Repeatability- and Reproducibility standard deviation the single values of the double determination were used.

Queried and documented were single results, recovery and the used testing method, information on the limit of quantification, the date of the analysis and general points to the method.

In case participants submitted several results for the same parameter obtained by different methods these results were evaluated with the same evaluation number with a letter as a suffix and indication of the related method.

Out of 10 participants, 9 participants submitted at least one result in time.

3. Evaluation

3.1 Consensus values from participants (Assigned value)

The robust mean of the submitted results was used as assigned value (X) („consensus value from participants“) providing a normal distribution. The calculation was done according to algorithm A as described in annex C of ISO 13528 [3].

The condition is that the majority of the participants' results show a normal distribution or are distributed unimodal and symmetrically. To this end, an examination of the distribution is carried out, inter alia, using the kernel density estimate [3, 12].

In case there are indications for sources of higher variability such as a bimodal distribution of results, a cause analysis is performed. Frequently different analytical methods may cause an anomaly in results' distribution. If this is the case, separate evaluations with own assigned values (X_{pti}) are made whenever possible.

The statistical evaluation is carried out for all the parameters for a minimum of 7 values are present.

The actual measurement results will be drafted. Individual results, which are outside the specified measurement range of the participating laboratory (for example with the result > 25 mg/kg or $< 2,5$ mg/kg) or the indicating "0" will not be considered for the statistic evaluation [3].

3.2 Robust standard deviation

For comparison to the target standard deviation σ_{pt} (standard deviation for proficiency assessment) a robust standard deviation (S^*) was calculated. The calculation was done according to algorithm A as described in annex C of ISO 13528 [3].

3.3 Repeatability standard deviation

The repeatability standard deviation S_r is based on the laboratory's standard deviation of (outlier free) individual participant results, each under repeatability conditions, that means analyses was performed on the same sample by the same operator using the same equipment in the same laboratory within a short time. It characterizes the mean deviation of the results within the laboratories [3] and is used by DLA as an indication of the homogeneity of the sample material.

In case single results from participants are available the calculation of the repeatability standard deviation S_r , also known as standard deviation within laboratories S_w , is performed by: [3, 4].

The relative repeatability standard deviation as a percentage of the mean value is indicated as coefficient of variation CV_r in the table of statistical characteristics in the results section in case single results from participants are available.

3.4 Reproducibility standard deviation

The reproducibility standard deviation S_R represents a inter-laboratory estimate of the standard deviation for the determination of each parameter on the bases of (outlier free) individual participant results. It takes into account both the repeatability standard deviation S_r and the within-laboratory standard deviation S_s . Reproducibility standard deviations of PT's may differ from reproducibility standard deviations of ring trials, because the participating laboratories of a PT generally use different internal conditions and methods for determining the measured values.

In the present evaluation, the specification of the reproducibility standard deviation, therefore, does not refer to a specific method, but characterizes approximately the comparability of results between the laboratories, assumed the effect of homogeneity and stability of the sample are negligible.

In case single results from participants are available the calculation of the reproducibility standard deviation S_R is performed by: [3, 4].

The relative reproducibility standard deviation as a percentage of the mean value is indicated as coefficient of variation CV_R in the table of statistical characteristics in the results section in case single results from participants are available. Its meaning is explained in more detail in 3.9.

3.5 Exclusion of results and outliers

Before statistical evaluation obvious blunders, such as those with incorrect units, decimal point errors, and results for a another proficiency test item can be removed from the data set [2]. Even if a result clearly deviates from the robust mean (e.g. factor >10) and has an influence on the robust statistics, a result can be excluded from statistical evaluation [3].

All results should be given at least with 2 significant digits. Specifying 3 significant digits is usually sufficient.

Results obtained by different analytical methods causing an increased variability and/or a bi- or multimodal distribution of results, are treated separately or could be excluded in case of too few numbers of results. For this results are checked by kernel density estimation [3, 12].

Results are identified as outliers by the use of robust statistics. If a value deviates from the robust mean by more than 3 times the robust standard deviation, it is classified as an outlier [3]. Detected outliers are stated for information only, when z-score are < -2 or > 2. Due to the use of robust statistics outliers are not excluded, provided that no other reasons are present [3].

3.6 Target standard deviation (for proficiency assessment)

The target standard deviation of the assigned value σ_{pt} (= standard deviation for proficiency assessment) can be determined according to the following methods.

If an acceptable quotient S^*/σ_{pt} is present, the target standard deviation of the general model by Horwitz is preferably used for the proficiency assessment. It is usually suitable for evaluation of interlaboratory studies, where different methods are applied by the participants. On the other hand the target standard deviation from the evaluation of precision data of a precision experiment is derived from collaborative studies with specified analytical methods.

In cases where both above-mentioned models are not suitable, the target standard deviation is determined based on values by perception, see under 3.6.3.

For information, the z-scores of both models are given in the evaluation, if available.

For the evaluation the target standard deviation from the general model of Horwitz (s. 3.6.1) was applied. For information, the target standard deviation of a precision experiment was given (ASU S64 Method: [18]-[20] and [22]), see 3.6.2/ Table 2).

3.6.1 General model (Horwitz)

Based on statistical characteristics obtained in numerous PTs for different parameters and methods Horwitz has derived a general model for estimating the reproducibility standard deviation σ_R [6]. Later the model was modified by Thompson for certain concentration ranges [10]. The reproducibility standard deviation σ_R can be applied as the relative target standard deviation σ_{pt} in % of the assigned values and calculated according to the following equations [3]. For this the assigned value x_{pt} is used for the concentration c .

Equations	Range of concentrations	corresponds to
$\sigma_R = 0,22c$	$c < 1,2 \times 10^{-7}$	$< 120 \mu\text{g/kg}$
$\sigma_R = 0,02c^{0,8495}$	$1,2 \times 10^{-7} \leq c \leq 0,138$	$\geq 120 \mu\text{g/kg}$
$\sigma_R = 0,01c^{0,5}$	$c > 0,138$	$> 13,8 \text{ g/100g}$

with c = mass content of analyte (as relative size, e.g. $1 \text{ mg/kg} = 1 \text{ ppm} = 10^{-6} \text{ kg/kg}$)

3.6.2 Precision experiment

Using the reproducibility standard deviation σ_R and the repeatability standard deviation σ_r of a precision experiment (collaborative trial or proficiency test) the target standard deviation σ_{pt} can be derived considering the number of replicate measurements m of participants in the present PT [3]:

$$\sigma_{pt} = \sqrt{\sigma_R^2 - \sigma_r^2 (m-1/m)}$$

The relative repeatability standard deviations (RSD_r) and relative reproducibility standard deviation (RSD_R) given in Table 2 were determined in ring tests using the indicated methods.

The resulting target standard deviations σ_{pt} , which were identified there, were used to evaluate the results and to provide additional information for the statistical data.

Table 2: relative repeatability standard deviations (RSD_r) and relative reproducibility standard deviations (RSD_R) according to selected evaluations of tests for precision and the resulting target standard deviation σ_{pt} [18 - 20, 22].

Parameter	Matrix	Mean	RSD_r	RSD_R	σ_{pt}	Method / Literature
Arsenic	Mushroom	0,07 mg/kg	18,6%	43%	40,9 %	ICP-MS [18]
Arsenic	Rice	0,95 mg/kg	8,12%	40,0%	39,6 % ¹	AAS-Hydride [22]
Cadmium	Mushroom	0,46 mg/kg	3,8%	6,9%	6,36 % ¹	ICP-MS [18]
Cadmium	Paprika	0,38 mg/kg	5,5%	20%	19,6 %	AAS [19]
Mercury	Mushroom	0,24 mg/kg	4,5%	16%	15,7 % ¹	ICP-MS [18]
Mercury	Trout	0,12 mg/kg	12,9%	23,0%	21,1 %	AAS [20]
Lead	Carrot	0,088 mg/kg	5,9%	12%	11,3 % ¹	ICP-MS [18]
Lead	Mushroom	1,5 mg/kg	15%	16%	12,0 %	ICP-MS [18]

¹ values used in the evaluation (on blue-grey), see section 4.

3.6.3 Value by perception

The target standard deviation for proficiency assessment can be set at a value that corresponds to the level of performance that the coordinator would wish laboratories to be able to achieve [3].

In the present PT, the target standard deviations according to 3.6.1 were considered suitable.

Table 3 shows selected characteristics of the participant results of the present PT compared to the results of previous years.

Tabelle 3: Characteristics of the present PT (on blue-gray) compared to the past PTs from 2014 (SD = Standard deviation, CV = variation coefficient)

Parameter	Matrix (powder)	rob. Mean	rob. SD (S*)	rel. SD (CV _{s*}) [%]	Quotient S*/ σ_{pt}	DLA-Report
Lead	Plant mixture	0,446 mg/kg	0,0517 mg/kg	11,6 %	0,64	32-2017
Lead	Plant mixture	1,48 mg/kg	0,152 mg/kg	10,3 %	0,68	33-2016
Lead	Plant mixture	3,40 mg/kg	0,414 mg/kg	12,2 %	0,92	23-2015
Lead	Plant mixture	7,50 mg/kg	1,83 mg/kg	24,4 %	2,07	22-2014
Cadmium	Plant mixture	0,464 mg/kg	0,0655 mg/kg	14,1 %	0,79	32-2017
Cadmium	Plant mixture	0,483 mg/kg	0,0373 mg/kg	7,72 %	0,43	33-2016
Cadmium	Plant mixture	1,81 mg/kg	0,121 mg/kg	6,68 %	0,45	23-2015
Cadmium	Plant mixture	0,128 mg/kg	0,029 mg/kg	22,7 %	1,0	22-2014
Arsenic	Plant mixture	0,378 mg/kg	0,0338 mg/kg	8,96 %	0,48	32-2017
Arsenic	Plant mixture	0,948 mg/kg	0,0549 mg/kg	5,79 %	0,36	33-2016
Arsenic	Plant mixture	0,946 mg/kg	0,182 mg/kg	19,2 %	1,2	23-2015
Arsenic	Plant mixture	0,797 mg/kg	0,156 mg/kg	19,6 %	1,2	22-2014
Mercury	Plant mixture	0,219 mg/kg	0,0367 mg/kg	16,8 %	0,83	32-2017
Mercury	Plant mixture	0,346 mg/kg	0,0843 mg/kg	24,4 %	1,3	33-2016
Mercury	Plant mixture	0,321 mg/kg	0,0376 mg/kg	11,7 %	0,62	23-2015
Mercury	Plant mixture	2,07 mg/kg	0,37 mg/kg	17,9 %	1,3	22-2014

3.7 z-Score

To assess the results of the participants the z-score is used. It indicates about which multiple of the target standard deviation (σ_{pt}) the result (x_i) of the participant is deviating from the assigned value (x_{pt}) [3].

Participants' z-scores are derived from:

$$z_i = \frac{(x_i - x_{pt})}{\sigma_{pt}}$$

The requirements for the analytical performance are generally considered as fulfilled if

$$-2 \leq z \leq 2 .$$

The z-score valid for the PT evaluation is designated z-score (σ_{pt}), while the value of z-score (Info) is for information only. The two z-scores are calculated using the different target standard deviations according to 3.6.

3.7.1 Warning and action signals

In accordance with the norm ISO 13528 it is recommended that a result that gives rise to a z-score above 3,0 or below -3,0, shall be considered to give an "action signal" [3]. Likewise, a z-score above 2,0 or below -2,0 shall be considered to give a "warning signal". A single "action signal", or "warning signal" in two successive PT-rounds, shall be taken as evidence that an anomaly has occurred which requires investigation. For example a fault isolation or a root cause analysis through the examination of transmission error or an error in the calculation, in the trueness and precision must be performed and if necessary appropriate corrective measures should be applied [3].

In the figures of z-scores DLA gives the limits of warning and action signals as yellow and red lines respectively. According to ISO 13528 the signals are valid only in case of a number of ≥ 10 results [3].

3.8 z'-Score

The z'-score can be used for the valuation of the results of the participants, in cases the standard uncertainty has to be considered (s. 3.8). The z'-score represents the relation of the deviation of the result (x) of the participant from the respective consensus value (X) to the square root of quadrat sum of the target standard deviation ($\hat{\sigma}$) and the standard uncertainty ($U_{x_{pt}}$) [3].

The calculation is performed by:

$$z'_i = \frac{x_i - x_{pt}}{\sqrt{\sigma_{pt}^2 + u_{(x_{pt})}^2}}$$

If carried out an evaluation of the results by means of z 'score, we have defined below the expression in the denominator as a target standard deviation σ_{pt} '.

The requirements for the analytical performance are generally considered as fulfilled if

$$-2 \leq z' \leq 2 .$$

For warning- and action-signals see 3.7.1.

3.9 Reproducibility coefficient of variation (CV)

The variation coefficient (CV) of the reproducibility (= *relative reproducibility standard deviation*) is calculated from the standard deviation and the mean as follows [4, 13]:

$$CV_R = \frac{S_R * 100}{\bar{x}}$$

In contrast to the standard deviation as a measure of the absolute variability the CV gives the relative variability within a data region. While a low CV, e.g. <5-10% can be taken as evidence for a homogeneous set of results, a CV of more than 50% indicates a "strong inhomogeneity of statistical mass", so that the suitability for certain applications such as the assessment of exceeded maximum levels or the performance evaluation of the participating laboratories possibly can not be done [3].

3.10 Quotient S^*/σ_{pt}

Following the Horrat-value the results of a proficiency-test (PT) can be considered convincing, if the quotient of robust standard deviation S^* and target standard deviation σ_{pt} does not exceed the value of 2. A value > 2 means an insufficient precision, i.e. the analytical method is too variable, or the variation between the test participants is higher than estimated. Thus the comparability of the results is not given [3].

3.11 Standard uncertainty

The consensus value has a standard uncertainty $U(X_{pt})$ that depends on the analytical method, differences between the analytical methods used, the test material, the number of participant laboratories (P) and perhaps on other factors. The standard uncertainty of the assigned value ($U(X_{pt})$) for this PT is calculated as follows [3]:

$$u_{(x_{pt})} = 1,25 \times \frac{s^*}{\sqrt{p}}$$

If $U(X_{pt}) \leq 0,3 \sigma_{pt}$ the standard uncertainty of the consensus value needs not to be included in the interpretation of the results of the PT [3]. A clear exceeded the value of 0.3 is an indication that the target standard deviation was possibly set too low for the standard uncertainty of the assigned value.

The quotient $U(X_{pt})/\sigma_{pt}$ is reported in the characteristics of the test.

4. Results

All following tables are anonymized. With the delivering of the evaluation-report the participants are informed about their individual evaluation-number.

In the first table the characteristics are listed:

Statistic Data
<i>Number of results</i>
<i>Number of outliers</i>
Mean
Median
Robust mean (X_{pt})
Robust standard deviation (S^*)
<i>Number with 2 replicates</i>
<i>repeatability standard deviation (S_r)</i>
Repeatability (Cv_r) in %
<i>reproducibility standard deviation (S_R)</i>
Reproducibility (Cv_R) in %
<i>Target range:</i>
Target standard deviation σ_{pt} or σ_{pt}'
Target standard deviation (for information)
lower limit of target range $(X_{pt} - 2\sigma_{pt})$ or $(X_{pt} - 2\sigma_{pt}')$ *
upper limit of target range $(X_{pt} + 2\sigma_{pt})$ or $(X_{pt} + 2\sigma_{pt}')$ *
<i>Quotient S^*/σ_{pt} or S^*/σ_{pt}'</i>
<i>Standard uncertainty $U(X_{pt})$</i>
<i>Quotient $U(X_{pt})/\sigma_{pt}$ or $U(X_{pt})/\sigma_{pt}'$</i>
<i>Results in the target range</i>
<i>Percent in the target range</i>

* Target range is calculated with z-score or z'-score

In the table below, the results of the participating laboratories are formatted in 3 valid digits**:

Auswerte- nummer	Parameter [Einheit/ Unit]	Abweichung	Z'-Score	z-Score	Hinweis
Evaluation number		Deviation	σ_{pt}'	(Info)	Remark

** In the documentation part, the results are given as they were transmitted by the participants.

4.1 Lead (mg/kg)

Vergleichsuntersuchung / Proficiency Test

Statistic Data	
Number of results	9
Number of outliers	1
Mean	0,513
Median	0,440
Robust Mean (X)	0,446
Robust standard deviation (S*)	0,0517
Number with 2 replicates	8
Repeatability SD (S_r)	0,0193
Repeatability (CV_r)	4,43%
Reproducibility SD (S_R)	0,0399
Reproducibility (CV_R)	9,16%
Target range:	
Target standard deviation σ_{pt}	0,0806
Target standard deviation (for Information)	0,0502
lower limit of target range	0,285
upper limit of target range	0,607
Quotient S^*/σ_{pt}	0,64
Standard uncertainty $U(X_{pt})$	0,0215
Quotient $U(X_{pt})/\sigma_p$	0,27
Results in the target range	8
Percent in the target range	89%

Comments:

For the valuation the target standard deviation of the general model according to Horwitz was applied. For information, the target standard deviation from a precision experiment (ASU § 64 LFGB L00.00-135) was given.

The distribution of the results showed a normal variability. The quotient S^*/σ_{pt} was well below 2,0. The coefficient of variation (with respect to the robust standard deviation, CV_{S^*}) is in the range of previous PTs (see 3.6.3). The comparability of results is given.

Repeatability- and reproducibility standard deviation are within the range of established values for the analytical methods used (see 3.6.2).

The quotient $U(X_{pt})/\sigma_p$ (0,27) is assessed to be low.

One result was not in the target range and is assessed as an outlier.

89% of the results were in the target area.




Abb. / Fig. 1: Ergebnisse Blei / Results lead




Abb. / Fig. 2:

Kerndichte-Schätzung der Ergebnisse
(mit $h = \sigma_{pt}$ von X_{pt})

Kernel density plot of results
(with $h = \sigma_{pt}$ of X_{pt})

Comment:

The kernel density shows a normal distribution of results with a side peak at 1,1 mg/kg, due to the result outside the target range (outlier).

Ergebnisse der Teilnehmer:**Results of Participants:**

Auswerte- nummer	Blei / lead [mg/kg]	Abweichung [mg/kg]	z-Score (σ_{pt})	z-Score (Info)	Hinweis
Evaluation number		Deviation [mg/kg]		(Info)	Remark
1	0,454	0,00731	0,091	0,15	
2	0,440	-0,00619	-0,077	-0,12	
3	0,398	-0,0482	-0,60	-1,0	
4	1,12	0,677	8,4	13	Ausreisser / Outlier
5	0,436	-0,0102	-0,13	-0,20	
6	0,440	-0,00619	-0,077	-0,12	
7	0,450	0,00381	0,047	0,08	
8	0,500	0,0538	0,67	1,1	
9	0,375	-0,0717	-0,89	-1,4	

**Abb. / Fig. 3:** Z-Scores Blei/ Lead

4.2 Cadmium (mg/kg)

Vergleichsuntersuchung / Proficiency Test

Statistic Data	
Number of results	9
Number of outliers	0
Mean	0,471
Median	0,460
Robust Mean (X)	0,464
Robust standard deviation (S*)	0,0655
Number with 2 replicates	9
Repeatability SD (S_r)	0,0225
Repeatability (CV_r)	4,78%
Reproducibility SD (S_R)	0,0751
Reproducibility (CV_R)	16,0%
<i>Target range:</i>	
Target standard deviation σ_{pt}	0,0833
Target standard deviation (for Information)	0,0295
lower limit of target range	0,297
upper limit of target range	0,631
Quotient S^*/σ_{pt}	0,79
Standard uncertainty $U(X_{pt})$	0,0273
Quotient $U(X_{pt})/\sigma_{pt}$	0,33
Results in the target range	9
Percent in the target range	100%

Comments:

For the valuation the target standard deviation of the general model according to Horwitz was applied. For information, the target standard deviation from a precision experiment (ASU § 64 LFGB L00.00-135) was given.

The distribution of the results showed a normal variability. The quotient S^*/σ_{pt} was well below 2,0. The coefficient of variation (with respect to the robust standard deviation, CV_{s^*}) is in the range of previous PTs (see 3.6.3). The comparability of results is given.

Repeatability- and reproducibility standard deviation are within the range of established values for the analytical methods used (see 3.6.2).

The quotient $U(X_{pt})/\sigma_p$ (0,33) is not increased.

100% of the results were in the target area.




Abb. / Fig. 4: Ergebnisse / Results Cadmium




Abb. / Fig. 5:

Kerndichte-Schätzung der Ergebnisse (mit $h = \sigma_{pt}$ von X_{pt})


Kernel density plot of results
(with $h = \sigma_{pt}$ of X_{pt})

Comment:

The kernel density shows a normal distribution of results.

Ergebnisse der Teilnehmer:**Results of Participants:**

Auswerte- nummer	Cadmium [mg/kg]	Abweichung [mg/kg]	z-Score (σ_{pt})	z-Score (Info)	Hinweis
Evaluation number		Deviation [mg/kg]			Remark
1	0,469	0,00453	0,054	0,15	
2	0,520	0,0560	0,67	1,9	
3	0,460	-0,00397	-0,048	-0,13	
4	0,629	0,165	2,0	5,6	
5	0,482	0,0175	0,21	0,59	
6	0,460	-0,00397	-0,048	-0,13	
7	0,460	-0,00397	-0,048	-0,13	
8	0,380	-0,0840	-1,0	-2,8	
9	0,384	-0,0805	-1,0	-2,7	

**Abb. / Fig. 6:** Z-Scores Cadmium

4.3 Arsenic (mg/kg)

Vergleichsuntersuchung / Proficiency Test

Statistic Data	
Number of results	8
Number of outliers	1
Mean	0,339
Median	0,380
Robust Mean (X)	0,378
Robust standard deviation (S*)	0,0338
Number with 2 replicates	7
Repeatability SD (S_r)	0,0103
Repeatability (CV_r)	2,67%
Reproducibility SD (S_R)	0,0251
Reproducibility (CV_R)	6,53%
Target range:	
Target standard deviation σ_{pt}	0,0700
Target standard deviation (for Information)	0,149
lower limit of target range	0,238
upper limit of target range	0,518
Quotient S^*/σ_{pt}	0,48
Standard uncertainty $U(X_{pt})$	0,0150
Quotient $U(X_{pt})/\sigma_p$	0,21
Results in the target range	7
Percent in the target range	88%

Comments:

For the valuation the target standard deviation of the general model according to Horwitz was applied. For information, the target standard deviation from a precision experiment (ASU § 64 LFGB L15.06-2) was given.

The distribution of the results showed a normal variability. The quotient S^*/σ_{pt} was well below 2,0. The coefficient of variation (with respect to the robust standard deviation, CV_{s^*}) is in the range of previous PTs (see 3.6.3). The comparability of results is given.

Repeatability- and reproducibility standard deviation are within the range of established values for the analytical methods used (see 3.6.2).

The quotient $U(X_{pt})/\sigma_p$ (0,21) is low.

One result was not in the target range and is assessed as an outlier.

88% of the results were in the target area.




Abb. / Fig. 7: Ergebnisse Arsen / Results arsenic




Abb. / Fig. 8:

Kerndichte-Schätzung der Ergebnisse
(mit $h = \sigma_{pt}$ von X_{pt})


Kernel density plot of results
(with $h = \sigma_{pt}$ of X_{pt})

Comment:

The kernel density shows a normal distribution of results with a side peak at 0,015 mg/kg, due to the result outside the target range (outlier).

Ergebnisse der Teilnehmer:**Results of Participants:**

Auswerte- nummer	Arsen/ Arsenic [mg/kg]	Abweichung [mg/kg]	z-Score (σ_{pt})	z-Score (Info)	Hinweis
Evaluation number		Deviation [mg/kg]		(Info)	Remark
1	0,399	0,0209	0,30	0,14	
2	0,410	0,0324	0,46	0,22	
3	0,410	0,0324	0,46	0,22	
4	0,0151	-0,363	-5,2	-2,4	Ausreisser / Outlier
5	0,366	-0,0121	-0,17	-0,081	
6	0,350	-0,0276	-0,39	-0,18	
7	0,390	0,0124	0,18	0,083	
8	0,370	-0,00761	-0,11	-0,051	

**Abb. / Fig. 9:** Z-Scores Arsen/ Arsenic

4.4 Mercury (mg/kg)

Vergleichsuntersuchung / Proficiency Test

Statistic Data	
Number of results	7
Number of outliers	0
Mean	0,219
Median	0,216
Robust Mean (X)	0,219
Robust standard deviation (S*)	0,0367
Number with 2 replicates	7
Repeatability SD (S_r)	0,0249
Repeatability (CV_r)	11,3%
Reproducibility SD (S_R)	0,0368
Reproducibility (CV_R)	16,8%
<i>Target range:</i>	
Target standard deviation σ_{pt}	0,0441
Target standard deviation (for Information)	0,0344
lower limit of target range	0,131
upper limit of target range	0,307
Quotient S^*/σ_{pt}	0,83
Standard uncertainty $U(X_{pt})$	0,0174
Quotient $U(X_{pt})/\sigma_{pt}$	0,39
Results in the target range	7
Percent in the target range	100%

Comments:

For the valuation the target standard deviation of the general model according to Horwitz was applied. For information, the target standard deviation from a precision experiment (ASU § 64 LFGB L00.00-135) was given.

The distribution of the results showed a normal variability. The quotient S^*/σ_{pt} was well below 2,0. The coefficient of variation (with respect to the robust standard deviation, CV_{s*}) is in the range of previous PTs (see 3.6.3). The comparability of results is given.

Repeatability- and reproducibility standard deviation are within the range of established values for the analytical methods used (see 3.6.2).

The quotient $U(X_{pt})/\sigma_p$ (0,39) lies at 0.39, just above 0,3 and is acceptable because of the other characteristics and the use of different analytical methods.

100% of the results were in the target area.




Abb. / Fig. 10: Ergebnisse Quecksilber / Results Mercury




Abb. / Fig. 11:

Kerndichte-Schätzung der Ergebnisse (mit $h = \sigma_{pt}$ von X_{pt})


Kernel density plot of results
(with $h = \sigma_{pt}$ of X_{pt})

Comment:

The kernel density shows a normal distribution of results.

Ergebnisse der Teilnehmer:**Results of Participants:**

Auswerte- nummer	Quecksilber/ Mercury [mg/kg]	Abweichung [mg/kg]	z-Score (σ_{pt})	z-Score (Info)	Hinweis
Evaluation number		Deviation [mg/kg]		(Info)	Remark
1	0,190	-0,0291	-0,66	-0,85	
2	0,240	0,0209	0,47	0,61	
3					
4	0,248	0,0290	0,66	0,84	
5	0,216	-0,00358	-0,081	-0,10	
6	0,260	0,0409	0,93	1,2	
7	0,170	-0,0491	-1,1	-1,4	
8	0,210	-0,00908	-0,21	-0,26	
9					

**Abb. / Fig. 12:** z-Scores Quecksilber/ Mercury

5. Documentation

5.1 Details by participants

5.1.1 Primary data

Note: Information given in German was translated by DLA to the best of our knowledge (without guarantee of correctness).

5.1.1.1 Lead

Teilnehmer	Proben-Nr. A	Proben-Nr. B	Datum d. Analyse	Ergebnis (Mittel)	Ergebnis A	Ergebnis B	Bestimmungs-grenze	Inkl. WF	Wiederfindungs-rate [%]
Participant	Sample No. A	Sample No. B	Date of analysis	Result (Mean)	Result A	Result B	Limit of quantification	Incl. RR	Recovery rate [%]
			day/month	mg/kg	mg/kg	mg/kg	mg/kg	yes/no	in %
1	32	40	23.10.17	0,4535	0,463	0,444	0,005	no	
2	30	69	24.10.17	0,44	0,43	0,44	0,0002		
3	9	63	19.10.17	0,398	0,383	0,414	0,0087	no	
4	11	61	24.10.17	1,1235	1,202	1,045	0,05		
5	62	10	17.10.17	0,436	0,439	0,433	0,005	no	
6	32	32	09.10.17	0,44	0,445	0,433	0,01		
7	46	26	11.10.17	0,45	0,482	0,422	0,05	no	94,5
8	31	41	04.10.17	0,50	0,51	0,49	0,1	no	100
9	12	60	24.10.17	0,3745	0,365	0,384		no	

5.1.1.2 Cadmium

Teilnehmer	Proben-Nr. A	Proben-Nr. B	Datum d. Analyse	Ergebnis (Mittel)	Ergebnis A	Ergebnis B	Bestimmungs-grenze	Inkl. WF	Wiederfindungsrate [%]
Participant	Sample No. A	Sample No. B	Date of analysis	Result (Mean)	Result A	Result B	Limit of quantification	Incl. RR	Recovery rate [%]
			day/month	mg/kg	mg/kg	mg/kg	mg/kg	yes/no	in %
1	32	40	23.10.17	0,4685	0,468	0,469	0,005	no	
2	30	69	24.10.17	0,52	0,52	0,51	0,00002		
3	9	63	19.10.17	0,46	0,46	0,46	0,00069	no	
4	11	61	24.10.17	0,6285	0,6754	0,5816	0,003		
5	62	10	17.10.17	0,4815	0,476	0,487	0,001	no	
6	32	32	09.10.17	0,46	0,458	0,461	0,01		
7	46	26	11.10.17	0,46	0,458	0,464	0,005	no	99,5
8	31	41	04.10.17	0,38	0,38	0,38	0,05	no	100
9	12	60	24.10.17	0,3835	0,387	0,380		no	

5.1.1.3 Arsenic

Teilnehmer	Proben-Nr. A	Proben-Nr. B	Datum d. Analyse	Ergebnis (Mittel)	Ergebnis A	Ergebnis B	Bestimmungsgrenze	Inkl. WF	Wiederfindungsrate [%]
Participant	Sample No. A	Sample No. B	Date of analysis	Result (Mean)	Result A	Result B	Limit of quantification	Incl. RR	Recovery rate [%]
			day/month	mg/kg	mg/kg	mg/kg	mg/kg	yes/no	in %
1	32	40	23.10.17	0,3985	0,394	0,403	0,005	no	
2	30	69	24.10.17	0,41	0,41	0,4	0,0013		
3	9	63	19.10.17	0,41	0,399	0,432	0,0044	no	
4	11	61	24.10.17	0,0151	0,0178	0,0124	0,009		
5	62	10	17.10.17	0,3655	0,359	0,372	0,013	no	
6	32	32	09.10.17	0,35	0,346	0,351	0,01		
7	46	26	11.10.17	0,39	0,391	0,388	0,01	no	98
8	31	41	04.10.17	0,37	0,37	0,37	0,1	no	100
9	12	60							

5.1.1.4 Mercury

Teilnehmer	Proben-Nr. A	Proben-Nr. B	Datum d. Analyse	Ergebnis (Mittel)	Ergebnis A	Ergebnis B	Bestimmungsgrenze	Inkl. WF	Wiederfindungsrate [%]
Participant	Sample No. A	Sample No. B	Date of analysis	Result (Mean)	Result A	Result B	Limit of quantification	Incl. RR	Recovery rate [%]
			day/month	mg/kg	mg/kg	mg/kg	mg/kg	yes/no	in %
1	32	40	23.10.17	0,190	0,196	0,184	0,005	no	
2	30	69	24.10.17	0,24	0,25	0,23	0,0001		
3	9	63							
4	11	61	24.10.17	0,24805	0,2283	0,2678	0,1		
5	62	10	17.10.17	0,2155	0,221	0,21	0,001	no	
6	32	32	09.10.17	0,26	0,28	0,24	0,01		
7	46	26	11.10.17	0,17	0,153	0,188	0,01	no	91,3
8	31	41	04.10.17	0,21	0,24	0,18	0,05	no	100
9	12	60							

5.1.2 Analytical methods

5.1.2.1 Lead

Teilnehmer	Methodenbeschreibung	Probenvorbereitung	Messmethode	Kalibrierung und Referenzmaterial	Wiederfindung mit gleicher Matrix	Methode akkreditiert	Sonstige Hinweise
Participant	Method description	Sample preparation	Measuring method	Calibration and reference material	Recovery with same matrix	Method accredited	Further remarks
					yes/no	yes/no	
1							
2	ICP-MS	Microwave pressure digestion (HNO ₃ /H ₂ O ₂)		external standard		no	
3		Pressure digestion with HNO ₃ , H ₂ O ₂	ICP-MS			no	
4	House method					yes	
5	AOAC 993.14	Microwave digestion		External calibration		yes	
6	EN ISO 170294-2	EN 15763	ICP-MS	(simple) linear		yes	
7	ASU §64 LFGB L.00.00-135	Microwave pressure digestion	ICP-MS	Liver powder	no	yes	
8	DIN EN ISO 17294-2 (E 29)	DIN EN 13805			yes	yes	
9	UNI EN 13806:2003	300mg, HNO ₃ + H ₂ O ₂ conc. - 200 °C, dil. 20 ml		Sigma Aldrich 16595 / batch: BCBR7883V	no	yes	

5.1.2.2 Cadmium

Teilnehmer	Methodenbe- schreibung	Probenvorbereitung	Messmethode	Kalibrierung und Referenzmaterial	Wiederfin- dung mit gleicher Ma- trix	Methode ak- kreditiert	Sonstige Hinweise
Participant	Method descrip- tion	Sample preparation	Measuring method	Calibration and reference material	Recovery with same matrix	Method ac- credited	Further remarks
					yes/no	yes/no	
1							
2	ICP-MS	Microwave pressure digestion (HNO ₃ /H ₂ O ₂)		external standard		no	
3						no	
4	House method					yes	
5	AOAC 993.14	Microwave digestion		External calibration		yes	
6	EN ISO 170294-2	EN 15763	ICP-MS	(simple) linear		yes	
7	ASU §64 LFGB L.00.00-135	Microwave pressure digestion	ICP-MS	Liver powder	no	yes	
8	DIN EN ISO 17294-2 (E 29)	DIN EN 13805			yes	yes	
9	UNI EN 13806:2003	300mg, HNO ₃ + H ₂ O ₂ conc. - 200 °C, dil. 20 ml		Sigma Aldrich 51994 / batch: BCBR8251V	no	yes	

5.1.2.3 Arsenic

Teilnehmer	Methodenbeschreibung	Probenvorbereitung	Messmethode	Kalibrierung und Referenzmaterial	Wiederfindung mit gleicher Matrix	Methode akkreditiert	Sonstige Hinweise
Participant	Method description	Sample preparation	Measuring method	Calibration and reference material	Recovery with same matrix	Method accredited	Further remarks
					yes/no	yes/no	
1							
2	ICP-MS	Microwave pressure digestion (HNO ₃ /H ₂ O ₂)		external standard		no	
3						no	
4	House method					yes	
5	AOAC 993.14	Microwave digestion		External calibration		yes	
6	EN ISO 170294-2	EN 15763	ICP-MS	(simple) linear		yes	
7	ASU §64 LFGB L.00.00-135	Microwave pressure digestion	ICP-MS	Liver powder	no	yes	
8	DIN EN ISO 17294-2 (E 29)	DIN EN 13805			yes	yes	
9							

5.1.2.4 Mercury

Teilnehmer	Methodenbe- schreibung	Probenvorberei- tung	Messmethode	Kalibrierung und Referenzmaterial	Wiederfindung mit gleicher Matrix	Methode akkredi- tiert	Sonstige Hinweise
Participant	Method descrip- tion	Sample preparati- on	Measuring method	Calibration and reference material	Recovery with same matrix	Method ac- credited	Further remarks
					yes/no	yes/no	
1							
2	ICP-MS	Microwave pressure digestion (HNO ₃ /H ₂ O ₂)		external standard		no	
3							
4	House method					yes	
5	EPA 7473	Direct mercury analysis		External calibration		yes	
6	EN ISO 170294-2	EN 15763	ICP-MS	(simple) linear		yes	
7	ASU §64 LFGB L.00.00-135	Microwave pressure digestion	ICP-MS	Liver powder	no	yes	
8	DIN EN ISO 12846 (E 12)	DIN EN 13805			yes	yes	
9							

5.2 Homogeneity

5.2.1 Homogeneity testing before PT

The mixture homogeneity before bottling was examined 10-fold by microtracer analysis.

Microtracer Homogeneity Test

DLA 32-2017

Weight whole sample	0,60	kg
Microtracer	FSS-rot lake	
Particle size	75 – 300	µm
Weight pro particle	2,0	µg
Addition of tracer	21,4	mg/kg

Result of analysis:

Sample	Weight [g]	Particle number	Particle [mg/kg]
1	9,73	98	20,1
2	8,07	81	20,1
3	10,08	101	20,0
4	9,40	95	20,2
5	9,42	98	20,8
6	9,41	111	23,6
7	9,18	103	22,4
8	7,98	90	22,6
9	9,03	97	21,5
10	9,57	96	20,1

Poisson distribution

Number of samples	10	
Degree of freedom	9	
Mean	97,1	Partikel
Standard deviation	5,99	Partikel
χ^2 (CHI-Quadrat)	3,32	
Probability	95	%
Recovery rate	99	%

Normal distribution

Number of sample	10	
Mean	21,1	mg/kg
Standard deviation	1,30	mg/kg
rel. Standardabweichung	6,2	%
Horwitz Standard deviation	10,1	%
HorRat Value	0,6	
Recovery rate	99	%

5.2.2 Comparison of sample number/test results and trend line

By comparison of the **increasing sample numbers** and the measurement results of iodine, the homogeneity of the chronological bottled PT item can be characterized with the help of the trend line function:

Cadmium			
Target standard deviation σ_{pt}	0,083		mg/kg
Sample numbers	9 - 69		
Total numbers of samples	18		
Slope	-0,00331		
Trend line range	0,503	-	0,444 mg/kg
Deviation trend line	0,473	±	0,0296 mg/kg
Percent of σ_{pt}	35,5	%	




Abb./Fig. 13:

Trendfunktion Probennummern vs. Ergebnisse
trend line function sample number vs. results

5.3 Sample cover letter: Information on the Proficiency Test (PT)

Before the PT, the participants are given the following information in the sample cover letter:

Information on the Proficiency Test (PT)

PT number	DLA 32-2017
PT name	Contaminated Food: Heavy Metals (Pb, Cd, Hg, As) in Root Vegetable Powder
Sample matrix*	Samples A + B: Root Vegetable Powder
Number of samples and sample amount	2 identical samples A + B, 5 g each.
Storage	Samples A + B: room temperature
Intentional use	Laboratory use only (quality control samples)
Parameter	quantitative: lead, cadmium, mercury and arsenic
Methods of analysis	Analytical methods are optional
Notes to analysis	The analysis of PT samples should be performed like a routine laboratory analysis. In general we recommend to homogenize a representative sample amount before analysis according to good laboratory practice, especially in case of low sample weights.
Result sheet	The results for sample A and B as well as the final results calculated as mean of the double determination (samples A and B) should be filled in the result submission file. The recovery rates, if carried out, has to be included in the calculation.
Units	mg/kg
Number of significant digits	at least 2
Further information	For information please specify: - Date of analysis - DLA-sample-numbers (for sample A and B) - Limit of detection - Assignment incl. Recovery - Recovery with the same matrix - Method is accredited
Result submission	The result submission file should be sent by e-mail to: pt@dla-lvu.de
Deadline	the latest October 27th 2017
Evaluation report	The evaluation report is expected to be completed 6 weeks after deadline of result submission and sent as PDF file by e-mail.
Coordinator and contact person of PT	Dr. Gerhard Wichmann

* Control of mixture homogeneity and qualitative testings are carried out by DLA. Testing of the content, homogeneity and stability of PT parameters is subcontracted by DLA.

6. Index of participant laboratories

Teilnehmer/ participant	Ort/ town	Land/ country
		Germany
		Italy
		USA
		Germany
		USA
		Germany
		Switzerland
		Germany
		Germany

[Die Adressdaten der Teilnehmer wurden für die allgemeine Veröffentlichung des Auswerte-Berichts nicht angegeben.]

[The address data of the participants were deleted for publication of the evaluation report.]

7. Index of literature

1. DIN EN ISO/IEC 17025:2005; Allgemeine Anforderungen an die Kompetenz von Prüf- und Kalibrierlaboratorien / General requirements for the competence of testing and calibration laboratories
2. DIN EN ISO/IEC 17043:2010; Konformitätsbewertung - Allgemeine Anforderungen an Eignungsprüfungen / Conformity assessment - General requirements for proficiency testing
3. ISO 13528:2015 & DIN ISO 13528:2009; Statistische Verfahren für Eignungsprüfungen durch Ringversuche / Statistical methods for use in proficiency testing by interlaboratory comparisons
4. ASU §64 LFGB: Planung und statistische Auswertung von Ringversuchen zur Methodenvalidierung / DIN ISO 5725 series part 1, 2 and 6 Accuracy (trueness and precision) of measurement methods and results
5. Verordnung / Regulation 882/2004/EU; Verordnung über amtliche Kontrollen zur Überprüfung der Einhaltung des Lebensmittel- und Futtermittelrechts sowie der Bestimmungen über Tiergesundheit und Tierschutz / Regulation on official controls performed to ensure the verification of compliance with feed and food law, animal health and animal welfare rules
6. Evaluation of analytical methods used for regulation of food and drugs; W. Horwitz; Analytical Chemistry, 54, 67-76 (1982)
7. The International Harmonised Protocol for the Proficiency Testing of Analytical Laboratories ; J.AOAC Int., 76(4), 926 - 940 (1993)
8. A Horwitz-like funktion describes precision in proficiency test; M. Thompson, P.J. Lowthian; Analyst, 120, 271-272 (1995)
9. Protocol for the design, conduct and interpretation of method performance studies; W. Horwitz; Pure & Applied Chemistry, 67, 331-343 (1995)
10. Recent trends in inter-laboratory precision at ppb and sub-ppb concentrations in relation to fitness for purpose criteria in proficiency testing; M. Thompson; Analyst, 125, 385-386 (2000)
11. The International Harmonised Protocol for the Proficiency Testing of Analytical Chemistry Laboratories; Pure Appl Chem, 78, 145 - 196 (2006)
12. AMC Kernel Density - Representing data distributions with kernel density estimates, amc technical brief, Editor M Thompson, Analytical Methods Committee, AMCTB No 4, Revised March 2006 and Excel Add-in Kernel.xla 1.0e by Royal Society of Chemistry
13. EURACHEM/CITAC Leitfaden, Ermittlung der Messunsicherheit bei analytischen Messungen (2003); Quantifying Uncertainty in Analytical Measurement (1999)
14. GMP+ Feed Certification scheme, Module: Feed Safety Assurance, chapter 5.7 Checking procedure for the process accuracy of compound feed with micro tracers in GMP+ BA2 Control of residues, Version: 1st of January 2015 GMP+ International B.V.
15. MTSE SOP No. 010.01 (2014): Quantitative measurement of mixing uniformity and carry-over in powder mixtures with the rotary detector technique, MTSE Micro Tracers Services Europe GmbH
16. HORWITZ EQUATION AS QUALITY BENCHMARK IN ISO/IEC 17025 TESTING LABORATORY, C. Rivera, R. Rodriguez, Pimentel 4104 -B; Col. Las Granjas. Chihuahua Chihuahua Mexico. C.P. 31160
17. AOAC Guidelines for Standard Method Performance Requirements (2016)
18. ASU §64 LFGB : L00.00-135; Bestimmung von Arsen, Cadmium, Quecksilber und Blei in Lebensmitteln (Jan. 2011)
19. ASU §64 LFGB : L00.00-19/3; Bestimmung von Blei, Cadmium, Chrom und Molybdän in Lebensmitteln (Juli 2004)
20. ASU §64 LFGB : L00.00-19/4; Bestimmung von Quecksilber in Lebensmitteln (Dezember 2003)
21. ASU §64 LFGB : L00.00-19/6; Bestimmung von Gesamtarsen in Lebensmitteln (Juli 2001)
22. ASU §64 LFGB: 15.06-2; Bestimmung von anorganischem Arsen in Reis mit Atomabsorptionsspektrometrie-Hydridtechnik nach Säureextraktion